



PRECISION DC POWER SOURCE

TECHNICAL DATA

MODEL 2005



POWER DESIGNS



0-20 V, 0-500 MA
A general purpose power supply with the accuracy of a precision calibrator

New high-stability, low-noise silicon semiconductor devices and advanced solid-state circuit technology are the basis for this unusually stable voltage regulator.

The improved performance of the Model 2005 permits dial readout of the output voltage to five places. Interpolation of the last place is provided by a potentiometer with a resolution of 10 microvolts.

Conventional, more complex chopper techniques for DC amplifier stabilization are avoided by maintaining critical amplifier stages and a specially processed zener diode voltage reference in a temperature-controlled oven.

The design simplicity results in a low-cost unit, half the size and weight of comparable instrumentation.

Design Features of the Model 2005P

- Calibrated decade voltage readout to four significant figures at outputs below 10 volts, to five significant figures above 10 volts. Interpolation of the last place is provided by a potentiometer with 10-microvolts resolution.
- Provisions for rear-panel zero calibration (may be used to offset lead drop during remote sensing).
- Adjustable current limiting.

- Self-restoring electronic overload and short-circuit protection.
- All silicon-semiconductor regulator system.
- Critical semiconductors and components maintained at constant ambient in temperature-controlled oven.
- Accurate remote programming at 1000 ohms-per-volt.
- Front and rear access to output terminals.
- 100 hour pre-aging of power supply before test and calibration. Individual calibration data furnished with each unit.
- Line and load circuits separately fused. Accessible at rear. Performance specifications based on anticipated ratings after 5 years service.
- "Controlled-Parameter" semiconductor program insures long life expectancy. Features "controlled avalanche" silicon rectifiers and power transistors, pre-aged zener voltage references and transistors, noise-testing techniques for establishing predictable device reliability, derating to 50% of rated voltage and current, etc.
- Modular package construction suitable for rack mounting. Single or dual mounting in 5¼" x 19" panel. See Cat. RPA-62 for rack panel adapters.

MODEL 2005



POWER DESIGNS

PRECISION DC POWER SOURCE



ELECTRICAL SPECIFICATIONS

OUTPUT - 0 to 20 VDC, continuously adjustable, at 0 to 500 MA.

INPUT - 105 to 125 V, 50 to 440 Hz, 40 W max. (including oven heater).

CALIBRATION AND ACCURACY - Two dual, concentric, decade switches provide dial readout to four places with an accuracy of $0.1\% \pm 1$ MV at outputs below 10 V. A 1-MV range, single-turn potentiometer permits interpolation of the last place. This potentiometer has a resolution of 10 μ V. A toggle switch extends the range of the dial readout from 10 to 20 V, without affecting resolution or accuracy (five-place readout). An illuminated indicator identifies the range in operation.

REGULATION - Output voltage change less than 100 μ V (at sense connection points) for line or load variations over the operating range.

RIPPLE AND NOISE - Less than 100 μ V peak-to-peak.

SOURCE IMPEDANCE - Less than 0.2 milliohm at DC, 0.04 ohm at 20 KHz, 0.5 ohm at 1 MHz.

RECOVERY TIME - Less than 10 μ s to return to within 250 μ V or 0.005% (whichever is greater) of the set voltage for a 10% to 100% step change in rated load; less than 40 μ s to return with 100 μ V.

OPERATING TEMPERATURE RANGE - 15 to 4 °C ambient; 0 to 60 °C ambient with reduced accuracy.

TEMPERATURE COEFFICIENT - DC output voltage change less than 0.001% or 50 μ V (whichever is greater) per °C over the range of + 15 to +45 °C; less than 0.002% or 100 μ V (whichever is greater) per °C from 0 to + 15 °C and +45 to +60 °C.

STABILITY - Better than 0.001% + 100 MV/8 hrs.; better than 1 MV/wk (at constant ambient temperature, load and line voltage, after warmup).

OUTPUT TERMINALS - Three front-panel binding posts and rear-panel barrier block. Either positive or negative output terminal may be grounded.

METER - Front-panel volt-ammeter permits monitoring output voltage or current with an accuracy of $\pm 2\%$.

REMOTE SENSING - Two terminals on rear-panel barrier block are provided for remote sensing of voltage at the load.

REMOTE PROGRAMMING - Rear terminals are provided for remote programming of the output voltage at 1000 ohms-per-volt. Accuracy of programming is better than 0.01% of value of resistance (including connecting leads).

CURRENT LIMITING - The output current can be limited to any value from 0 to 500 MA by a front-panel control. A push-button permits easy adjustment without the need for shorting the output terminals.

MECHANICAL SPECIFICATIONS

DIMENSIONS - 8 3/8" x 4 3/4" x 8-15/16"

WEIGHT - 9 pounds

FINISH - Brushed anodized aluminum panel with etched black lettering. Dust cover finished in blue enamel; chassis and bottom plate, golden indite.



Pair of Model 2005 Units shown in Type RRB-12 Rack Panel Adapter.

PRICE
\$325.00
FOB Westbury

PRINTED IN U.S.A.

POWER DESIGNS, INC.
1700 SHAMES DRIVE • WESTBURY, N.Y.
516 EDgewood 3-6200 TWX: 510-222-6561

POWER DESIGNS PACIFIC, INC.
3281 JUNIPERO SERRA • PALO ALTO, CALIFORNIA
415-321-6111 TWX: 910-373-1251

PRECISION DC POWER SOURCE MODEL 2005

SECTION 1

GENERAL DESCRIPTION

1-1. DESCRIPTION

The Model 2005 is a precision DC power source designed to supply an extremely stable 0 to 20 volt, 0 to 500 milliamperes output. The instrument combines the accuracy of a precision calibrator with the power capability of a general-purpose regulated supply.

Two dual, concentric decade switches provide a digital readout of the selected output voltage to within 0.1% +1 millivolt of the selected value. A 1-millivolt range, single-turn potentiometer permits interpolation of the last place. This potentiometer has a resolution of 10 microvolts. A toggle switch selects the range of the dial readout; either from 0 to 10 volts, or from 10 to 20 volts.

The output voltage of the supply may be remotely programmed with the same accuracy, using an external resistance. The supply also includes provisions for remote sensing of the output voltage at the load.

Compact and light, the power source is self-contained in a portable housing designed for bench use. The modular construction of the Model 2005 makes it suitable for rack mounting. Panel adapters are available for mounting one or two units in a standard 19-inch rack having a panel height of 5 1/4 inches.

1-2. ELECTRICAL SPECIFICATIONS

Refer to Table 1 for a complete list of electrical specifications.

1-3. MECHANICAL SPECIFICATIONS

a. Dimensions : 8-3/8 inches wide by 4-3/4 inches high by 8-15/16 inches deep behind the front panel.

b. Weight : 9 pounds.

c. Finish : The panel is finished in brushed aluminum and has etched black lettering. The housing is finished in blue-gray vinyl enamel. The chassis and bottom plate are gold iridite.

TABLE 1. ELECTRICAL SPECIFICATIONS

| <u>Parameter</u> | <u>Value</u> |
|----------------------------|---|
| Output | 0 to 20 volts DC, continuously adjustable, 0 to 500 milliamperes |
| Input | 105 to 125 volts, 50 to 440 Hz, 40 watts (nominal) |
| Regulation | DC voltage change less than 100 microvolts for line variations of $\pm 10\%$ or load variations of 100% (at sense lead connection points) |
| Ripple and Noise | Less than 100 microvolts peak-to-peak |
| Source Impedance | Less than 0.2 milliohm at DC, 0.04 ohm at 20 KHz, 0.5 ohm at 1 MHz |
| Recovery Time | Less than 10 microseconds to return to within 250 microvolts or 0.005% (whichever is greater) of the set voltage for a step change in rated load (1 microsecond rise time) of 10% to 100% or 100% to 10%; less than 40 micro- seconds to return to within 100 microvolts |
| Stability | Better than 0.001% +100 microvolts per 8 hours; better than 1 millivolt per week (at constant line, load and ambient temperature after warm-up). |
| Temperature Coefficient | DC output voltage change less than 0.001% or 50 microvolts (whichever is greater) per $^{\circ}\text{C}$ over the range of $+15^{\circ}\text{C}$ to $+45^{\circ}\text{C}$, less than 0.002% or 100 microvolts (whichever is greater) per $^{\circ}\text{C}$ from 0°C to $+15^{\circ}\text{C}$ and from $+45^{\circ}\text{C}$ to $+60^{\circ}\text{C}$ |
| Calibration Accuracy | Better than 0.1% +1 millivolt |
| Current Limiting | 0 to 500 milliamperes continuously adjustable by a front panel control. A front panel push button permits easy adjustment without shorting the output terminals. |
| Output Terminals | |
| Front Panel | Three insulated binding posts for positive output, negative output, and chassis ground |
| Rear Panel | Screw terminals on a molded barrier block for positive output, negative output, chassis ground, remote voltage programming and remote sensing. |
| Remote Sensing | Two terminals are provided on a rear panel barrier block for remote sensing of the voltage at the load. |



2005

Parameter

Specifications

| | |
|--------------------|--|
| Remote Programming | Rear panel barrier block terminals are provided for remote programming of the output voltage. The ratio of the programming resistance to the output voltage is 1000 ohms per volt. The programming accuracy is better than 0.01% of the resistance value, including the resistance of the programming leads. |
| Metering | Front panel volt-ammeter permits monitoring output voltage or current with an accuracy of ±2% of full scale. NOTE: This accuracy is considerably less than that of the power source. |
| Circuit Protection | The AC line and DC load circuits are separately fused. The fuses are accessible at the rear of the unit. |
| Indicator Lamps | |
| AC Lamp | Lights when AC toggle switch is set to ON and power is applied to the unit. |
| OVEN Lamp | Lights when AC input is applied and oven heater is energized (operates even when AC toggle switch is turned off). Lamp cycles on and off as oven maintains constant temperature environment for critical components. |
| +10 V Lamp | Lights when RANGE switch is set to 10-20V position to indicate that output is 10 volts plus the dial readout. |

2005

SECTION 2

INSTALLATION AND OPERATION

2-1. INSTALLATION

a. Laboratory Bench . The Model 2005 is a portable unit designed for bench use. No preliminary processing or special unpacking procedures are required. The power source is ready for operation as shipped from the factory.

b. Rack Mounting . Panel adapters are available for mounting one or two units in a standard 19-inch relay rack. Hardware kits are provided with each panel adapter. The power source is fastened to the rear surface of the adapter with the black anodized rivnuts in each corner of its front panel. If the rubber bumper feet interfere with the assembly of the equipment to the rack, they may be removed by disassembling the bottom plate of the power source and unscrewing them.

2-2. OPERATION

a. Operating Procedures:

- (1) Ensure that the power source is turned off and the shorting links are connected between the following rear-panel terminals: DC+ and S+; S+ and RV; and S and DC-.
- (2) Set the CURRENT LIMIT ADJ control fully clockwise.
- (3) Connect the AC line cord to a source of 105 to 125 volts, 50 to 440 Hz.

NOTE

The OVEN indicator will normally light as soon as line voltage is applied to the unit, even when the power source is turned off. The indicator will remain lighted until the oven reaches operating temperature (approximately 10 minutes). The indicator will cycle on and off as the oven maintains a constant-temperature environment for critical circuit components.

- (4) Read the information in paragraphs b through e to connect for remote sensing, series operation, remote programming of the output voltage, or current limiting.
- (5) Set the AC switch to ON.

- (6) Set the output voltage to the desired value, using the front panel dials. Set the VERNIER control to zero if interpolation between 1 millivolt steps is not desired.
- (7) Connect the DC+ and DC- terminals (on the front or rear panel) across the load. If desired, connect the positive or negative output terminal to ground.

b. Sensing . The regulator circuit maintains the potential between the sense leads (S+ and S-) at the set output voltage. When these leads are connected to the positive and negative output terminals, the power source is connected for local sensing. When the sense leads are connected to the load, the source is connected for remote sensing. Remote sensing is used when an appreciable voltage drop is anticipated in the leads connecting the positive and negative output terminals to the load. The Model 2005 is connected for local sensing when shipped from the factory. For remote sensing, proceed as follows:

- (1) Remove the shorting links from between the rear panel DC+ and S+ terminals and from between the DC- and S- terminals.
- (2) Connect the DC+ and DC- leads across the load.
- (3) Connect the S+ and S- leads to the positive and negative sides of the load, respectively. Run the sense leads as a tightly twisted, shielded pair. Connect the shield to the G (chassis ground) terminal to minimize output ripple
- (4) Turn on the Model 2005.

c. Series Operation . As many as four Model 2005 units may be connected in series to provide up to 80 volts. Connect the positive DC output terminal of one supply to the negative output terminal of the next, in the same manner as connecting batteries in series. The ground terminals on all units may be left floating or they may be tied together and connected to either the most positive or most negative output terminal.

To insure optimum voltage regulation, disconnect the shorting links between all S+ and DC+ output terminals except those at the most positive potential. Then connect jumper wires between each S+ terminal and the S- terminal on the next more positive power source. In this way the voltage drops in the leads connecting the power sources will be compensated for by the regulator circuits of the individual units. For remote sensing, proceed as directed above, except connect the most positive and most negative S+ and S- leads across the load.

d. Remote Voltage Programming . The output voltage can be programmed remotely by an external fixed or variable resistance. Proceed as follows:

- (1) Turn off the power source, set all output voltage controls to zero and set the RANGE switch to 0 -10V.
- (2) Remove the shorting link from between the rear panel RV and S+ terminals.

- (3) Select a programming resistance by multiplying the desired output voltage by 1000 (the programming constant is 1000 ohms per volt). A constant current of 1 milli-ampere will flow through this resistance, and its wattage rating should be chosen to minimize drift due to heating.
- (4) Connect the external programming resistance between the RV and S+ terminals using twisted, shielded wire. Connect the shield to the G (chassis ground) terminal to minimize output ripple.
- (5) Turn on the power source.

CAUTION

IF THE REMOTE PROGRAMMING CONNECTIONS ARE OPENED WHILE THE SUPPLY IS OPERATING, THE OUTPUT VOLTAGE WILL RISE SLIGHTLY ABOVE THE SET VALUE. WHEN A SWITCH IS USED TO SELECT RESISTORS FOR OUTPUT VOLTAGE PROGRAMMING, IT SHOULD HAVE SHORTING TYPE CONTACTS TO AVOID VOLTAGE SPINES.

e. Current Limiting : The maximum output may be limited to a value below 500 milliamperes as follows:

- (1) Turn on the power source and set the meter switch to MA.
- (2) Depress the CURRENT LIMIT SET button and adjust the CURRENT LIMIT ADJ control until the meter indicates the maximum desired output current. Release the CURRENT LIMIT SET button.

SECTION 3

PRINCIPLES OF OPERATION

3-1. GENERAL

The Model 2005 is a highly accurate, series regulated, DC voltage source. Basically, it consists of a full-wave rectifier circuit, a series regulator circuit and a current limiting circuit.

The series regulator circuit is essentially an electronically variable resistance interposed between the unregulated source and the load. The resistance value is controlled by an amplifier which compares the source output voltage with a reference voltage. The amplifier adjusts the series resistance to reduce the error signal to zero.

The reference voltage is generated by an internal auxiliary power source. The error signal resulting from the voltage comparison is amplified and applied through a driver stage to the series transistor to vary its effective resistance.

3-2. FULL-WAVE RECTIFIER OPERATION

The full-wave rectifier consists of diodes CR1 and CR2. Its output is applied through fuse F2 to series regulator transistor Q1. The output of the full-wave rectifier is filtered by capacitor C1.

3-3. SERIES REGULATOR OPERATION

The series regulator circuit consists of differential amplifier Q5, amplifiers Q4, Q6 and Q7; driver Q2 and series regulator Q1. The voltage reference for the differential amplifier is zener diode CR13. A constant current is maintained through CR13 by transistors Q9 and Q10 and zener diode CR11. These components are powered by an auxiliary 20-volt supply.

The constant voltage across CR13 is impressed across the base-emitter junction of one half of Q5. Potentiometer R12 is adjusted to bias the base-emitter junction of the input half of Q5 to the same potential. Zener diode CR10 provides a constant collector voltage for the input section of the differential amplifier. A constant current for this zener diode is provided by CR22 and Q10, which are powered by an auxiliary 20 volt supply.

The differential amplifier, its voltage reference and the transistors which maintain a constant current through the voltage reference are located on oven board assembly A1. The oven maintains these components in a constant-temperature environment to provide highly stable operation.

The input to the differential amplifier is applied from a voltage divider across the supply output. Any change in output voltage changes the bias on the differential amplifier and, consequently, changes the collector current on the output half of this stage. This



changes the drive on amplifiers Q4, Q6, and Q7. The changed output of Q7 changes the drive of Q2, and therefore of series regulator Q1. This change in drive on Q1 is in the correct direction to oppose any change in the supply output voltage.

For example, if the output voltage tends to increase, the forward bias on the input stage of the differential amplifier increases. This reduces the collector current of the output half of this stage, reducing the drive of amplifiers Q4, Q6, and Q7. The reduced collector current of Q7 lowers the forward bias of driver Q2. The reduced collector current of Q2 reduces the forward bias of series regulator Q1, increasing its effective resistance. The increased resistance of Q1 increases the voltage drop across it, reducing the output voltage.

3-4. CURRENT-LIMITING CIRCUIT

The current-limiting circuit consists of transistor Q8, diode CR18, and their associated components. This stage is connected across the auxiliary 20-volt supply. The current through resistors R21 and R25 through R27 sets the normal bias on this stage. Potentiometer R27 sets the range of CURRENT LIMIT ADJ R25. This potentiometer is adjusted to forward bias transistor Q8 at the current-limiting point.

When the output load demand exceeds the value set by potentiometer R25, transistor Q8 conducts heavily. The collector current of Q8, flowing through R29, forward biases diode CR18. When diode CR18 conducts, it reduces the forward bias of amplifier Q4. This reduces the drive of the series regulator transistor, lowering the output voltage.

Any further increase in load demand further reduces the bias on Q4, further reducing the output voltage. In this manner, the circuit will maintain the load current at the set value for loads down to a short circuit. When the output current demand is reduced, the circuit conditions reverse and the voltage regulating circuits regain control of the output.



2005
SECTION 4
MAINTENANCE

4-1. GENERAL

Under normal conditions, no special maintenance of the Model 2005 is required. If servicing is necessary, the information in this section should be read thoroughly before starting repair or calibration.

4-2. ADJUSTMENT AND CALIBRATION

Make the following adjustments whenever a component is replaced or periodic recalibration is scheduled:

a. Preliminary Meter Adjustment:

(i) Mechanically zero the meter using the adjustment screw on the front panel.

NOTE

Before completing any meter adjustment, lightly tap the meter face. This will overcome any pivot friction and insure proper calibration.

(2) Set the CURRENT LIMIT ADJ control fully clockwise.

(3) *Connect a 40—ohm, 10—watt resistor, in series with a standard ammeter, across the output of the supply.*

(4) Set the AC switch to ON and adjust the output voltage controls for a convenient reading, approximately ampere on a standard ammeter.

(5) Set the meter switch to MA.

(6) Adjust potentiometer R42 (on the board behind the meter) until the panel meter reading agrees with that of the standard ammeter.

b. Zero Voltage Calibration:

(1) Connect a high precision voltmeter across the output of the supply.

(2) Set the RANGE switch to the 0-10V position and the output voltage controls for 0.000 volt output.

(3) Set the AC switch to ON and observe the voltmeter.

(4) *If the voltmeter does not indicate 0.000 volt ± 1 MV, adjust CALIBRATE potentiometer R39 (on the rear panel) for this reading. If the adjustment is beyond the range of R39, proceed as follows:*

- (a) Set CALIBRATE potentiometer R39 to the middle of its range.
- (b) Disconnect any jumper wires connected across resistors R9, R13, R40 and/or R41.
- (c) Connect jumper wires, in various combinations, across resistors R9, R13, R40, and R41 until the output is within 2.5 MV of zero. Solder the jumper wires in these positions.
- (d) Adjust CALIBRATE potentiometer R39 for an output of 0.000 volt ± 1 MV.

c. 20-Volt Adjustment:

NOTE

Make this adjustment only after the zero voltage calibration.

(1) Set the RANGE switch to 10–20V and adjust the supply output to 20.000 volts. Set the VERNIER control fully counterclockwise.

(2) Check that the dots on the VERNIER control and front panel are aligned. A setscrew is located in the VERNIER control for any necessary adjustment.

(3) Set the VERNIER control to 0.

(4) Connect a high precision voltmeter across the output of the supply.

(5) Set the AC switch to ON and observe the voltmeter.

(6) If necessary, adjust potentiometer R12 (on the amplifier board) until the voltmeter reads 20.000 ± 0.007 volts.

d. CURRENT LIMIT ADJ Range Adjustment:

(1) Set the power supply output voltage to 20.000 volts.

(2) Set the meter switch to MA.

(3) Set potentiometer R27 (on the amplifier board) to the center of its range.

(4) Adjust CURRENT LIMIT ADJ potentiometer R25 through its entire range while depressing CURRENT LIMIT SET pushbutton S3.

(5) Adjust potentiometer R27 until the CURRENT LIMIT ADJ potentiometer varies the output current from 0 to 500 MA over its entire range.



- (6) Set the CURRENT LIMIT ADJ potentiometer to the desired maximum output current, or set it fully clockwise.

4-3. TROUBLE SYMPTOMS AND SUGGESTED REMEDIES

a. Circuit faults can be isolated most rapidly by measuring the voltage and resistance. Use the data given on the schematic diagram in the Appendix as a first stop in servicing the supply.

CAUTION

WHEN UNSOLDERING SEMICONDUCTORS FOR TEST, USE A HEAT SINK TO PREVENT THERMAL DAMAGE. A LONG NOSE PLIERS BETWEEN THE SEMICONDUCTOR AND THE SOLDER JUNCTION IS ADEQUATE. NEVER OPERATE THE POWER SUPPLY WITH ANY LEADS DISCONNECTED OR SEMICONDUCTORS REMOVED. OPERATING POTENTIALS IN THE DC AMPLIFIER MAY CHANGE RADICALLY WHEN A COMPONENT IS REMOVED OR DISCONNECTED.

b. Power Supply Does Not Go On : If the OVEN and AC lamps do not light, check the AC fuse. If the fuse blows repeatedly, check the oven circuit and diodes CR1, CR2 and CR4 through CR7. Use an ohmmeter to take a resistance reading across each diode. Then, reverse the meter leads and take another reading. If one reading is not at least five times greater than the other, the diode is defective. If one diode in any pair is defective, replace both. A short circuit in one will produce high surge currents in the other, which can result in junction damage.

c. No DC Output Voltage: If both the AC and OVEN lamps light, but no output voltage is available, insure that the CURRENT LIMIT ADJ control is not turned fully counterclockwise. Set the meter switch to MA and increase the output voltage. If no current is indicated, check the DC fuse and input capacitor C1. If current is present when the output-voltage controls are adjusted, check safety diode CR12 and for incorrect programming or sensing connections. Diode CR12 is connected in the opposite polarity to the DC output voltage. If the reverse current flow is greater than 1 ampere, this diode may weld, placing a permanent short circuit across the supply output. Normal operation can be restored by replacing the diode (located on the amplifier board).

d. Regulator Failure : Check for correct potentials on amplifier transistors, voltage reference, etc. If any voltage appears incorrect, disconnect AC power and make a rapid check for defective transistors. This can be done without removing the transistors from the circuit. Use an ohmmeter set to its low resistance scale (R x1) and measure the forward and reverse resistances at the collector-base and base-emitter junctions. A resistance ratio of less than 5 to 1 indicates that the transistor is defective. Carefully remove it and check it on a transistor checker.

2005

CAUTION

THIS TEST IS NOT RECOMMENDED FOR HIGH FREQUENCY OR LOW CURRENT DEVICES IN OTHER INSTRUMENTATION, AS CURRENTS FROM SOME OHMMETERS MAY BE SUFFICIENT TO DAMAGE SMALL SEMICONDUCTOR JUNCTIONS.

To test a component located in the oven:

- (1) Unplug the line cord and remove the cover from the power supply.
- (2) Loosen the three screws which secure the oven cover.
- (3) Rotate the cover counterclockwise and pull it away from the oven.
- (4) Remove the two screws which secure the oven cap; then remove the cap.
- (5) Reach into the oven and extract the oven board.

NOTE

To test the oven board while the unit is operating, remove it from its socket and insert a test adapter (Vector Electronic Corp. Type P-9-N-S, or equal) in its place. The board can then be plugged into the adapter.

- (6) To reassemble the oven, replace the board and cap. Secure the cap in position with two screws. Slide the oven cover down until the screws slide into the slots in the cover. Turn the cover clockwise and tighten the three screws.

e. Poor Regulation, High Ripple: No specific check can be suggested since failure to regulate within specifications may be caused by any of the components in the supply. Make a point-to-point voltage and resistance check. Check all capacitors for open circuits and all electrolytic capacitors for excessive leakage. Make stage-gain measurements by changing the output load current and noting the change in base current of each amplifier stage. Use low resistance milliammeters and microammeters to avoid upsetting the regulator. The open-loop current gain of the regulator should be more than 106 from the base current of the input differential amplifier to the collector current of the series regulator.